

SKLAR STAINLESS STEEL SURGICAL INSTRUMENTS RECOMMENDED CARE & CLEANING INSTRUCTIONS

CAUTION: The following instructions are for all Sklar stainless steel surgical instruments. Read instructions prior to use. Improper operation and care or use for purposes other than that intended can lead to premature wearing out of the surgical instruments. Always process new instruments prior to initial use (see step 4).

1. PRE-TREATMENT

General Blood and body fluids can cause pitting on instruments and if left to dry, can be difficult to remove. In order to achieve successful decontamination, disinfection and sterilization, always wipe or rinse instruments immediately after use. If rinsing is not immediately available, pre-treat instruments with a neutral pH/enzymatic solution at point of service.

2. RINSING

General Immediately after surgery, remove organic material by rinsing soiled instruments under cold running water. Never process Steel, Magnesium, Aluminum or Zinc based materials with Stainless Steel. Always wear protective apparel as a standard precaution. Refer to OSHA and AORN standards for recommended precautions.

3. CLEANING

General Transport instruments to decontamination processing/cleaning area. Always keep soiled instruments covered during transport to prevent exposure to blood borne pathogens or other potentially infectious organisms. Before beginning the cleaning process, ensure that instruments have been thoroughly rinsed with copious amounts of cool running water. Separate instruments with dissimilar metals. Several methods of cleaning are available. Improper cleaning methods can result in damage to instruments or equipment and limit the warranty. See also AAMI TIR12.

Soak An enzymatic cleaning bath or neutral pH detergent effectively breaks down organic material from instruments when fully submerged for 10 minutes. Take care to prevent sharp tips (scissors, knives, osteotomes, etc.) from touching. Do not mix dissimilar metals. Thoroughly rinse instruments with cool running tap water (use distilled or demineralized water if possible) to remove solution(s). Change solutions as directed.

**Manual
Cleaning**

Always wash instruments in a manner that provides proper decontamination.

1. Mix a neutral pH detergent / enzymatic solution with luke-warm water following the manufacturer's mixing instructions and immerse instruments if possible. Highly acidic or highly alkaline pH detergents are not recommended for use on Sklar instruments.
2. Use a soft nylon brush to manually scrub instruments, concentrating on hinged areas, crevices and other difficult to clean locations. Limit use of stainless steel brushes to serrated areas, bone files or burs.
3. Brush delicate instruments carefully, separating them from general instruments whenever possible.
4. Prevent scratching by not allowing sharp tips (scissors, knives, osteotomes, etc.) from touching.
5. Visibly check instruments to ensure surfaces are clean and free from damage, stains and bioburden.
6. Check instruments for proper function and condition: smooth blade closure and opening, proper jaw alignment, working hinges, and proper locking ratchets.
7. Thoroughly rinse instruments using running water (distilled or demineralized water is recommended), paying close attention to hinged areas, box locks and moving parts to ensure they are rinsed thoroughly and no debris remains.
8. If to be stored, use a clean, lint-free cloth to dry instruments.
9. Lubricate instruments prior to packaging or sterilization.
10. Always store in a clean, dry environment

**Ultrasonic
Cleaning**

Mechanical cleaning of surgical instruments is the preferred cleaning method as it efficiently removes soil and provides consistent washing and rinsing parameters.

1. Fully submerge all instruments in an open position to effectively clean hinges, box locks and other moving parts. Prevent sharp tips (scissors, knives, osteotomes, etc.) from touching to avoid scratching. Do not mix dissimilar metals. Use distilled or demineralized water if possible.
2. Follow ultrasonic cleaner manufacturer's operating instructions.
3. Rinse instruments with water to remove cleaning solution(s).
4. Always lubricate instruments prior to sterilization. Regular lubrication is essential to ensure the life of instruments.

Caution: Processing needle holders and forceps with the ratchet in a closed position may crack box locks and hinges.

Automatic Washer Disinfectors	Mechanical cleaning of surgical instruments is the preferred cleaning method as it efficiently removes soil and provides consistent washing and rinsing parameters. Follow manufacturer's operating instructions. Ensure instruments are lubricated after the final rinse cycle and before sterilization.
-------------------------------	---

4. STERILIZATION (Follow these steps for processing new instruments prior to initial use)

General	All blood, body fluids and tissue should be completely removed from instruments prior to sterilization. Separate dissimilar metals prior to sterilization.
Lubrication	Lubrication is key to preserving the proper function of your instruments. Lubricate all hinged instruments that have metal-to-metal contact at the screw or box lock. A non-silicone, water-soluble lubricant is recommended. Do not rinse. Do not use industrial oils or lubricants.
Autoclaving	1. Process instruments individually or in sets. 2. Protect sharp tips. Place heavy instruments on the bottom of sets. 3. Always process all instruments in the open position. Instruments locked during sterilization can develop cracked hinges or other problems resulting from heat expansion. 4. Autoclave instruments according to AAMI ST79 standards. Sklar stainless steel instruments have been validated for the following steam sterilization cycles (wrapped configuration): Pre-Vacuum Steam, 4min. @ 132°C/270°F with 20 min. Dry Time. Gravity Displacement Steam, 15 min. @ 132°C/270°F with 15- 45 min. Dry Time. Note: Make sure autoclave chambers are cleaned regularly and as recommended by the manufacturer.
Chemical/ Cold Sterilization	Fully immerse clean, dry instruments in solution. Most chemical/cold sterilization solutions render instruments sterile after a minimum of 10 hours. Follow sterilant manufacturer's instructions. Caution: Prolonged chemical action can be detrimental to instruments. Chemical / cold sterilization solutions are not recommended for use on Tungsten Carbide instruments.

Note: The re-processor is responsible for ensuring that its processes are performed adhering to procedures, and using the requisite equipment, materials, and personnel in its facility to achieve the desired result. Routine monitoring and validation of the process is necessary to ensure the integrity of the process and result. Likewise, any deviation by the user from the instructions provided must be rigorously evaluated for effectiveness and to identify potential adverse consequences.

To properly care for and maintain the life of your instruments, Sklar recommends its line of cleaning agents.

SKLAR CLEANING PRODUCT LINE

P/N	Description	Application
10-2722	Sklar Spray-Zyme 22oz. Bottles, 12/cs	Pre-Treatment
10-2724	Sklar Spray-Zyme 1 Gallon Bottles, 4/cs	
P/N	Description	Application
10-2777	Sklar Enzymatic Detergent 1 Gallon Bottles, 4/cs	Cleaning Soak
P/N	Description	Application
10-2701	Sklar Kleen Low Foam 1 Gallon Bottles, 4/cs	Cleaning Soak
10-1607	Sklar Kleen Powder 5lb. Container, 1/ea	Manual Cleaning
10-1630	Sklar Kleen Powder 3.5lb. Pitcher, 4/cs	Automatic Washer Disinfectant
10-1612	Sklar Kleen Liquid 8oz. Bottles, 6/cs	
10-1613	Sklar Kleen Liquid 1 Gallon Bottles, 4/cs	
P/N	Description	Application
10-1635	Sklar Instru-Guard Lube 1 Gallon Concentrate, 4/cs	Lubrication
10-1636	Sklar Instru-Guard Lube 8oz. Spray Bottles, 12/cs	
P/N	Description	Application
10-1629	Sklar Polish Corrosive Stain Remover, 8 oz. Jar, 6/cs	
10-2738	Sklar Sheen Spray Foam Cleaner and Conditioner, 22 oz. Bottles, 12/cs	

Note: It is the responsibility of the reprocessor to ensure that the reprocessing, as actually performed using equipment, materials and personnel in the reprocessing facility, achieves the desired result. This requires validation and routine monitoring of the process. Likewise any deviation by the user from the instructions provided must be properly evaluated for effectiveness and potential adverse consequences.

For intraocular ophthalmic instruments, refer to the Ophthalmic Cleaning and Sterilization Addendum (Document ID: MKF970.21 Ophthalmic-Addendum Rev.00)."